Machine Learning role in clinical decision-making: Neuro-rehabilitation video game

Shabnam Sadeghi Esfahlani, Hassan Shirvani, Javaid Butt, Iraj Mirzaee, Karim Sadeghi Esfahlani

PII: S0957-4174(22)00554-1 DOI: <https://doi.org/10.1016/j.eswa.2022.117165> Reference: ESWA 117165 To appear in: *Expert Systems With Applications*

Received date : 29 October 2020 Revised date : 1 June 2021 Accepted date : 31 March 2022

Please cite this article as: S.S. Esfahlani, H. Shirvani, J. Butt et al., Machine Learning role in clinical decision-making: Neuro-rehabilitation video game. *Expert Systems With Applications* (2022), doi: [https://doi.org/10.1016/j.eswa.2022.117165.](https://doi.org/10.1016/j.eswa.2022.117165)

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2022 Published by Elsevier Ltd.

Click here to view linked References

Journal Pre-proof

05741_Manuscript with Highlights.pdf

Machine Learning Role in Clinical Decision-Making: Neuro-Rehabilitation Video Game

Shabnam Sadeghi Esfahlani^{a,c,∗}, Hassan Shirvani^{a,c}, Javaid Butt^{a,c}, Iraj Mirzaee^{b,c}, Karim Sadeghi Esfahlania,c

Journal Pre-proof *^aAnglia Ruskin University, Engineering and the Built Environment, CM1 1SQ, Bishop Hall Lane, Essex, UK. ^bUrmia University of Technology, Urmia, Iran ^cEmail addresses:shabnam.sadeghi-esfahlani@aru.ac.uk (Shabnam Sadeghi Esfahlani), Hassan.Shirvani@aru.ac.uk (Hassan Shirvani), Javaid.Butt@aru.ac.uk (Javaid Butt), i.mirzaee@urmia.ac.ir (Iraj Mirzaee), esfahlani.k@gmail.com (Karim Sadeghi Esfahlani).*

∗Author correspondents. *Email address:* shabnam.sadeghi-esfahlani@aru.ac.uk (Shabnam Sadeghi Esfahlani) *Preprint submitted to Elsevier May 31, 2021*

ract

Tradity, we investigated the potential use of Machine Learning algorithms (ML) to predict

utcome of home-based neuro-rehabilitation video game intervention and its advantage in

utcome of home-based neuro-rehabilita Abstract In this study, we investigated the potential use of Machine Learning algorithms (ML) to predict the outcome of home-based neuro-rehabilitation video game intervention and its advantage in supporting clinical decision-making. We adopted Support Vector Machines (SVM) and K-Nearest Neighbours (KNN) to develop multidimensional functions (multi-variable Kernel functions) since both algorithms were considered significant and active analysis agents for prediction and classification. Supervised SVM and KNN algorithms were trained using the upper extremity (arm, forearm, and hand) joints' kinematic data and hand gestures of participants while interacting with the developed video games. Data collected from healthy and Multiple sclerosis (MS) participants were compared and used to develop the predictive algorithm. Pre- and post-rehabilitation data of MS subjects were investigated and used to assess the subject's functional improvements following the program. Bayesian optimization, Sigmoid, polynomial, and Gaussian Radial Basis functions were utilized

for training and predicting outcomes. The results showed that the first two kernel regressions had the best performance regarding predictability and cross-validation loss. KNN's prediction accuracy was exceeded by 91.7% versus SVM, which was 88.0%. The effectiveness of the rehabilitation program was assessed through Spatiotemporal control and motor assessment scale presenting 40% improvement. Our findings suggest that ML has a great potential to be used for decision-making in neuro-rehabilitation programs.

Keywords: Machine Learning algorithm, Rehabilitation, Serious Video Game, Support Vector Machine, K-Nearest Neighbours

1. Introduction

- Multiple sclerosis (MS) is the most prevalent autoimmune disease affecting the nervous sys-
- ³ tem and motor control (Steinman 1996). Motor control is a mechanism by which individuals

Preprint submitted to Elsevier May 31, 2021

 use their cognition to stimulate and coordinate the muscles and limbs involved in the admin- istration of a motor ability (Benedict et al. 2005). Based on numerous studies, continuous in- patient or outpatient rehabilitation could lead to intensification inactivity and overall ability to τ engage in society with optimal improvement and functionality (Jonsdottir et al. 2018). According to (Kwakkel et al. 2019) UE, recovery probably occurs through an aggregate of inevitable and learning-dependent manners, including reestablishing the quality of movement and learning ways to use their residual capacity in the most practical way to accomplish a task.

 MS casualties often do not adhere to recommended exercise by clinicians due to lack of mo- tivation (Giusti et al. 2006). However, there is substantial evidence of the feasibility and effec- tiveness of using serious games for rehabilitation (Esfahlani, Thompson, et al. 2018; Jonsdottir et al. 2018; Tannous 2018), which encourage participants' retention and incentive to practice.

 A study by (Jonsdottir et al. 2018) showed the clinically significant enhancements in MS peo- ple's arm function following interaction with a serious game for rehabilitation. (Bettger and ¹⁷ Stineman 2007) also suggested that regular exercises and automatic assessment of the upper ex- tremity through home-based rehabilitation could be advantageous for people with motor control impairment and MS condition.

ion of a motor ability (Renedict ct al. 2005). Based on numerous studies, continuous in-
to to outpatient rehabilitation could lead to intensification inactivity and overall ability to
ge in society with optimal improveme To facilitate a home-based rehabilitation program for MS casualties, we developed mini video ²¹ games using the principle of modified constraint-induced movement theory (mCIMT) and mir-₂₂ ror image therapy. The game's difficulty level was adjusted using the Monte Carlo Tree Search algorithm, examined in our previous studies (Esfahlani, Muresan, et al. 2018; Esfahlani, Butt, and Shirvani 2019). ML has received less attention in supporting clinical decision-making and predicting rehabilitation of upper extremity (UE) outcomes. Machine Learning (ML) algorithms were integrated to automate the process of analyzing 3D kinematics of the UE movements and ₂₇ classify its patterns based on data obtained from sensors (Microsoft Kinect, Leap Motion, and Myo armband devices). KNN and SVM ML algorithms were utilized due to their high accuracy and good theoretical guarantee of overfitting.

 Support Vector Machines (SVMs) and k-Nearest-Neighbours (KNNs) are the non-probabilistic 31 ML algorithms, which represent different approaches to learning as active analysis agents for

³² prediction and classification. These algorithms function based on a statistical learning procedure 33 demonstrated to have high accuracy (Stetter et al. 2020; Derie et al. 2020). The algorithm could ³⁴ predict the players' arm function following the rehabilitation program and classify the arm and hand pattern and gestures. We obtained a view-invariant representation of the 3D positions and orientation of the skeleton joints of UE, an arm's muscle actions, and hand gestures. Partici-³⁷ pant's with MS condition was distinguished from healthy subjects with the algorithm followed ³⁸ by predicting the motor function improvement following the intervention.

2. Background

onstrated to have high accuracy (Stetter et al. 2020; Deric et al. 2020). The algorithm could
the players' ann function following the rehabilitation program and classify the argn and
pattern and gestures. We obtained a vie ⁴⁰ ML has been successfully applied to several applications; ranging from face identification (Heisele, Ho, and Poggio 2001) to text categorization (Joachims 1998), and pattern recognition and classification problems (Levinger et al. 2007; Hosomi et al. 2012). The application of SVM to classify gait patterns of Knee osteoarthritis was conducted by (Levinger et al. 2007) to in- vestigate whether ML can assess gait improvement following knee replacement surgery. (Zhang et al. 2011) investigated optimal model selection for posture recognition through a supervised classification and training of a multiclass ML. They classified nine everyday postures from a 47 belt-worn smartphone's accelerometer data. (Lau, Tong, and H. Zhu 2009) explored the use of ML to classify different walking conditions for hemiparetic subjects. The participants walked ⁴⁹ in five different conditions in that two portable sensor units, comprising an Accelerometer and Gyroscope, were attached to the lower limb on the shank and foot segments to measure the kine- matic data. Their results showed that the SVM classification method could be applied as a tool for pathological gait analysis, pattern recognition, and activity monitoring during the rehabili- tation of daily exercises. They also suggested that the performance of an SVM was superior to other ML methods. (Begg, Palaniswami, and Owen 2005) integrated SVM to perform an auto- mated recognition of gait changes among young and senior participants. It was used to identify the aging influence on gait patterns and locomotor balance with the advantage of early identi- fication of at-risk gait and monitoring the progress of treatment outcomes. They recommended that SVMs function as an efficient gait classifier to recognize young and elderly gait patterns.

 (Fleury, Vacher, and Noury 2010) used a health smart home in that a wearable kinematic sensor was utilized to collect postural transitions using pattern recognition and walk periods frequency 61 analysis. The data collected from various sensors used to classify each temporal frame using SVM. (Long et al. 2016) utilized an online SVM optimized by particle swarm optimization to ⁶³ identify different locomotion modes to realize a smooth and automatic transition. Their experi-⁶⁴ mental results show the effectiveness of the SVM algorithm with high accuracy. (Oskoei, Hu, et ⁶⁵ al. 2008) showed that the SVM is computationally an efficient algorithm for classification using electromyogram inputs.

3. Materials and Methods

 The task in ML involves separating data into training and testing sets. Each instance in the training set contains one target value and several attributes. ML produces a model based on the τ_0 training data, which predicts the target values of the test data given only the test data attributes (Boser, Guyon, and V. N. Vapnik 1992). In this study, ML Matlab Toolbox was used to map the input into a high-dimensional space.

3.1. Support Vector Machines (SVM)

itilized to collect postural transitions using pattern recognition and walk periods frequency
sis. The data collected from various sensors used to classify each temporal frame using
1. (Long et al. 2016) utilized un online SVM constructs an optimal hyperplane as a decision surface to maximize the margin of separation between the two classes. An iterative training algorithm must construct an optimal hyperplane, which is managed to minimize an error function. Based on the error function, SVM models can be classified into four distinct types (Joachims 1998), categorized as classification and regression models. The algorithm finds optimal locations of the decision surface by using a set of mathematical functions involving classification and regression models (Cortes and V. Vapnik 1995). Training for the algorithm has two phases; (I) Remodel predictors (input data) to 81 a high-dimensional feature space where the data is never explicitly transformed into the feature ⁸² space. (II) Solve an optimization problem to fit an optimal hyperplane to classify the transformed ⁸³ features into two classes. The number of support vectors determines the number of transformed ⁸⁴ features, and the process of rearranging the objects is known as mapping or transformation.

VM require the solution for equations listed in Table 1. **Table shows two elassification and**
 Starting the solution for equations listed in Table 1. Table shows two elassification and
 Experimental Willis separating Given a training set of instance-label pairs (x_i, y_i) , i = 1,...,l where $x_i \in R^n$ and $y \in \{1, -1\}^l$, 86 the SVM require the solution for equations listed in Table 1. Table shows two classification and **87** two-regression types; C-SVM and V -SVM, and ϵ -SVM. $C \in [0, \infty]$, $V \in [0, 1]$ and $\epsilon \in [0, 1]$ ⁸⁸ are regulation parameters that supports implementing a penalty on the misclassifications that are ⁸⁹ performed while separating the classes. It helps in improving the accuracy of the output. The ⁹⁰ regularization parameters control the trade-off between the slack variable penalty (misclassifica-⁹¹ tions) and the width of the margin. Small value makes the constraints be ignored, which leads to 92 a large margin. Large value allows the constraints difficult to be ignored, which leads to a small ϕ margin. ϕ is the kernel function which maps χ_i to $\phi(\chi_i)$. ζ is the slack variable that allows ⁹⁴ regression errors to exist up to the value of ζ_i and ζ_i^* , yet still satisfy the required conditions. ⁹⁵ **W** ∈ Rⁿ is inversely proportional to margin, hence to maximize the margin, we will have to min-96 **imize W**, $b \in R$. γ K function estimates the functional dependence of the dependent variable γ on a set of inde-

98 pendent variables χ (ibid.). Kernel functions are distinguished mainly based on the localization ⁹⁹ and boundary between different classes to determine finite response across the entire range of ¹⁰⁰ features. Linear, polynomial, radial basis function (RBF), and sigmoid kernel functions were 101 adopted, which accompanied by Bayesian optimization to map data, illustrated in Eqn.1. Dot ¹⁰² product in the equation acts as a transformer to map input data points to the higher dimensional 103 feature space, and γ is kernel function's adjustable parameter (Joachims 1998).

$$
K(X_i, X_j) = \phi(X_i) \cdot \phi(X_j)
$$
\n
$$
exp(-\gamma |X_i - X_j|^2) \quad Gaussian RBF
$$
\n
$$
tanh(\gamma X_i \cdot X_j + C)
$$
\n
$$
Sigmoid
$$
\n(1)

104

¹⁰⁵ *3.2. K-Nearest Neighbour*

¹⁰⁶ Inputs in KNN consist of k closest training examples in the feature space. The output depends ¹⁰⁷ on whether KNN is qualified for classification or regression. In regression, the output is the

Table 1: SVM algorithm with the error functions and constraints.

¹⁰⁸ object's property value that is the average of the KNNs values. KNN regression's output is the ¹⁰⁹ object's property value that an average of the values of CNN's. In KNN classification, the output ¹¹⁰ is a class membership where a majority vote of the neighbors classifies objects. KNN classifier or ¹¹¹ weighted nearest neighbor classifier could be perceived as assigning the K-Nearest Neighbours ¹¹² a weight $\frac{1}{k}$. Given a X set of n points and a distance function, KNN search finds the k most 113 adjacent points in X to set of points Y . The choice of K is essential and needs to be selected 114 carefully, i.e., if K is too large or small, some of the neighbors used to make prediction will ¹¹⁵ no longer be similar to the foreseen one, which will bias the prediction (Hastie, Tibshirani, and 116 Friedman 2009). In the study, an optimal K is selected empirically, examining cross-validation 117 procedure on the training set (ibid.).

118 Given the covariate vector of a new observation, x_0 , the goal is to predict its response, y_0 . For every observation x_i in the training set, let $s_i = s(x_0, x_i)$ be its similarity to x_0 . Then the $s_{(1)} \geq s_{(2)} \geq \cdots \geq s_{(n)}$. Similarly, if $s_j = s_{(k)}$, where x_j is the kth most similar observation in training set to x_0 then the set of KNNs of x_0 ; $N(\chi_0, K)$, could be described as all observations whose similarities to χ_0 are at least $S(K)$;

123 $N(x_0, K) = \{x_i : s_i \geq s_{(K)}\}.$

$$
\hat{\rho} = \frac{\sum_{x_i \in N(x_0, k)} y_i}{|N(x_0, K)|}
$$
\n(2)

124 KNN algorithm estimates the probability of $y_0 = 1$, $|N(x_0, K)|$ is the number of items con- 125 tained in the set $N(\chi_0, K)$. It generally equals K but may exceed it depending on how ties are 126 treated. The response is then predicted to be one if $\hat{p} \geq c$, where c is a prespecified thresh-127 old parameter. The overall error rate in KNN decreases as K increases and levels off at around $K = 20$. In a few cases, we can see that the overall error rate commences increasing again as K, 129 increasing further; thus, we chose K to be 20.

¹³⁰ *3.3. Video Game* & *Modified Constraint-Induced Movement Therapy*

 $\hat{p} = \frac{\sum_{N} eM(x_0, t_0)Y_l}{|N(x_0, K)|}$

algorithm estimates the probability of $y_0 = 1$, $|N(x_0, K)|$ is the number of items con-

ai an the set $N(x_0, K)$. It generally equals K but may exceed it depending on how ties are
 Constraint-Induced Movement Therapy (CIMT) is a physical rehabilitation strategy that uses 132 operant training techniques applied in the context of rehabilitation medicine (Gert Kwakkel et al. 2015), developed by (Taub, Crago, and Uswatte 1998). Its principle is to continue stretching mo- tor capacity gradually beyond an attained achievement level. CIMT emphasizes massed practice with the affected upper limb by restraining the less affected limb and training the affected one by shaping movements. CIMT is exhaustive, possibly resulting in non-compliance with the proto- col. It devotes six hours or more of therapy while constraining the intact arm for 90% of waking hours per day throughout two weeks (Gert Kwakkel et al. 2015). Thus, a modified version of it - mCIMT has been formed by (Page et al. 2008) to overcome such complexity. The mCIMT intends to overcome learned non-use in chronic hemiparesis, which is the behavioral conquest of purposive movement of the more affected UE in daily living exercises. Two to three weeks of mCIMT for stroke patients' rehabilitation have shown significant improvements in the spon- taneous use of the paretic limb in the live setting, in comparison with placebo control therapy or usual and customary care (Wolf et al. 2006). (Mark et al. 2008) also suggested that slowly progressive MS conditions could take advantage of mCIMT and achieve promising progress. Since chronic UE hemiparesis occurs in MS casualties (Cowan, Ormerod, and Rudge 1990), we hypothesized that such patients could manifest learned non-use and favorably respond to the therapy. Therefore, in this study, the mCIMT principle was paired with mirror image therapy

ition of behavioral techniques known as shaping. Shaping involves matching the difficulty
sks performed to the improvements the patients nake and providing encouraging feedfleatly
sks performed to the improvements the pati to acquire mini video games. It was developed based upon repetitive task practice and the ap- plication of behavioral techniques known as shaping. Shaping involves matching the difficulty of tasks performed to the improvements the patients make and providing encouraging feedback immediately after any gain in function (Corbetta et al. 2015). Shaping in conventional mCIMT is determined by therapists based on individual movement deficits at specific skeleton joints (joint movements) that have the most potential for improvement according to a therapist's judgment (Taub, Uswatte, et al. 2006). Whereas in our design, the process and decision-making are con- ducted automatically by the algorithm. The progression of movement tasks is made systemati- cally, quantified, and parametric way on personalized tasks for a patient. Duration of intervention in conventional mCIMT could vary from [2-10] weeks. Furthermore, the treatment time could be modified from thirty minutes to three hours per session (Yen et al. 2005). Thus, therapy sessions are designed as follows:

- ¹⁶¹ 1. Therapy sessions of one hour, five days a week for ten weeks.
- 2. Restraining the use of a non-paretic upper limb to promote the use of the more impaired limb during sessions.
- 3. Adherence-enhancing behavioral methods designed to transfer the gains obtained in the clinical setting to the patients' real-world environment.

166 4. Each task was practiced for at least five minutes before starting the formal session.

- 167 Table 2 lists the mini-video-game tasks developed for this study with the devices used to interact
- 168 with the therapy/game. Further explanations on devices and the game setting are followed.
- 1. forearm to the table, participants' forearm reaches the highlighted areas presented in the virtual environment.
- 2. Forearm to the virtual button, and hold it in that position for three seconds, while the visual feedback and timer are activated.
- 3. Extend an elbow to press the virtual button.
- 4. Hand to the table and press the virtual button at the same time.
- 5. Hand to the virtual button and hold it for three seconds.

Table 2: Task-specific use of devices and computer supervised mini-video games for the rehabilitation.

¹⁷⁶ 6. Collect virtual boxes and place them on top of each other.

177 7. Lift virtual boxes to 10 cm (identified by highlighted areas in the game).

¹⁷⁸ 8. Pull the virtual rubber sling and release it to smash/hit virtual boxes.

179 9. Grab virtual objects that are spawned randomly in the space to collect them.

¹⁸⁰ 10. Lift a virtual pencil and put it on the virtual box.

¹⁸¹ 11. Stack virtual boxes (up to four boxes),

- ¹⁸² 12. Flip the virtual cards.
- ¹⁸³ 13. Grip/hold an object firmly,
- 184 14. Turn a key in a virtual lock.
- ¹⁸⁵ 15. Reach virtual fruits and place them in the basket.

¹⁸⁶ The video game in this work provides participants with graphical feedback, positive rewarding ¹⁸⁷ and scoring system to encourage the continuation of tasks and retention. The player acts as a

¹⁸⁸ controller of the game in that body actions are transferred into the 3D environment using the 189 mirroring effect via Kinect and Leap Motion.

3.3.1. Myo Armband

191 Myo armband by Thalmic Labs is a low-cost consumer-grade EMG device integrating an ARM Cortex-M4 based microcontroller unit, a set of eight dry EMG electrodes, a nine-axis inertial measurement unit (IMU), and a Bluetooth Low Energy (BLE) module (Phinyomark, N Khushaba, and Scheme 2018). The Myo is non-intrusive, as the electrodes allow users to slip 195 the bracelet on and off (Côté-Allard et al. 2018; Esfahlani, Thompson, et al. 2018). **EMG and** ¹⁹⁶ IMU data, were collected for hand gestures. The optimal position of the armband placement is determined by conducting various examinations and placed on the widest part of the forearm (right or left). For optimal results and a strong connection to the arm muscles, the Myo armband 199 was warmed up for five minutes (Phinyomark, N Khushaba, and Scheme 2018).

3.3.2. Microsoft Kinect V2

flyo amband by Thalmic Labs is a low-cost consumer-grade EMG device integrating an

If Cortex-M4 based microcontroller unit, a set of eight dry EMG electrodes, a mne-axis

al measurement unit (IMU), and a Bluetooth Low En Kinect v2 has an Infrared (IR) camera with 512 **×** 424 pixels. Its RGB camera resolution is 202 1920 **×** 1080 pixels. It has a field of view (FoV) of 70 **×** 60° and a frame rate of 30Hz. Kinect 203 was located on top of the screen within [1.5-2.0] meters from the player. It was elevated 45[°] in front of the subject (tilted toward the subject). Kinect was used to transfer the players' body joints to the game environment. Skeleton joints were also used for classification and regression analysis. To measure the dynamic range of motion of the UE joints, we employed a sequential kinematic model based on the joint coordinate system (JCS), which was proposed by the Standardization and Terminology Committee (STC) of the International Society of Biomechanics (ISB) (Wu et al. 2005). The coordinates were defined where the coronal axis of the patient is the X-axis, the sagittal axis is the Y-axis, the vertical axis is the Z-axis, and the idle position is the origin of the 211 joint coordinates (Butler et al. 2010).

3.3.3. Leap Motion

 Leap Motion peripheral uses two monochromatic IR cameras and three IR LEDs. It observes a roughly hemispherical area to a distance of about one meter. Leap Motion was located on the desk using a projective interaction mode within 0.4 to 0.6 meters below the player's hands. Figure 1-a, Figure 1-b, and Figure 1-c are the screenshots of task number six using Leap Motion.

Figure 1: (a), (b) and (c) 6th Video game task with virtual hands following the physical hands using Leap Motion. (d), (e) and (f) Rubber sling to knock down virtual boxes (right).

217 The player should reach virtual boxes to interact with them. In this scenario, both hands are involved in completing the task where mirrored therapy is used in a 3D environment. The boxes ²¹⁹ must be picked up and placed on the pads matching the cube's color. At the more advanced level, the player should stack boxes on top of each other. Figure 1-d, Figure 1-e, and Figure 1-f is the screenshot of task number eight from mini-game, according to Table 2.

ovel in completing the task where mirrored therapy is used in a 3D environment. The hoses
be picked up and placed on the pads matching the cube's color. At the more advanced level,
hyer should stack boxes on top of each ot Figure 1-e, and Figure 1-f both show that the player should only use the left hand to interact with the virtual ball and the elastic sling. Figure 2 illustrates the location of virtual fruits, task number $_{224}$ fifteen in Table 2, spawned in the 3D environment. Monte Carlo Tree Search algorithm was 225 used to spawn objects relative to the position of the player's body joints and their performance. According to the figure, the player only can interact with objects using the right hand. Left-hand could not be used to collect things to promote the use of the affected hand/arm. Figure 2 displays various planes in which the objects are spawned; abduction/adduction, forward flexion/extension, external/internal rotation and horizontal flexion/extension.

4. Calculation

4.1. Data Pre-processing and Feature Extraction

 Myo armband, Microsoft Kinect, and Leap Motion devices enable the player to act as a con- troller. Kinect collects the player's body joints orientation, and position, Leap Motion transfers fingers and hand movements into the game environment.

 EMG signals are formed via the superimposition of individual action potentials generated by ₂₃₆ irregular discharges of functioning motor units in a muscle fibre. Thus, amplitude and frequency both represent the level of activity of motor units in the fibre. EMG raw signals are transformed into a classification set of features following (feature extraction). There were three types of fea- tures in a different domain; Time, Frequency, and Time-Frequency distribution (Nazmi et al. 2016).

 Data pre-processing method of baseline correlation was performed using a 4th-order Butterworth bandpass filter with different cut-off frequencies to improve the signal-to-noise ratio (SNR) of the signals. Savitzky-Golay filter was integrated to remove noise from the IMU signals. The 50

(c) External/Internal-Rotation around shoulder/torso height (d) External/Internal-Rotation around shoulder/torso height

Figure 2: $15th$ video-game scenario of Table 2 with the location of the fruit which is spawned in the virtual environment relative to the player's performance (right hand).

 Hz notch power line noise removed using a 3 dB passband, amplitude normalization followed by wavelet method to localize both time and frequency segments (Phinyomark, N Khushaba, and Scheme 2018). EMG signals were recorded at 200 Hz with a high-pass 20 Hz cut-off and low-pass of 500 Hz cut-off at the time domain and frequency of 24 and 2.

 The EMG raw signal features were extracted for each type of hand movement from the denoised data. Waveform length time-domain feature extraction method was used to present the charac- teristics of signals for hand movements due to its high rate of accuracy and stability to changes in segmentation method (Oskoei, Hu, et al. 2008).

avel and the localize both time and frequency segments (Phinyomark, N. Khushaha,
Scheme 2018). EMG signals were recorded at 200 Hz with a high-pass 20 Hz cut-off and
Scheme 2018). EMG signals were recorded at 200 Hz with Four hand gestures were segmented using KNN and dynamic time warping algorithms. Each segment consisted of 500 data points taken from the armband. The armband's angular and linear movements were obtained from its Intertial Measurement Units (IMU: Gyroscope, Accelerome- ter, and Magnetometer) sensor. The Gyroscope determines the roll/pitch, and the magnetometer measures the yaw in various directions. After collecting data at the local sampling frequency, linear interpolations were performed to create a constant sampling interval while maintaining time and frequency domain signal integrity to achieve an identical sampling frequency. The sig- nals were averaged over a hundred and twenty readings to manage high-dimensional data sets efficiently. Acceleration creates a force that is captured by the force detection mechanism of the ²⁶¹ accelerometer, which measures acceleration indirectly through a force applied to its axis. Linear acceleration was measured in $(G=9.81 \frac{m}{s^2})$ along each axes considering static (gravity) and dynamic (sudden starts/stops) acceleration. The Gyroscope's readings were collected in $\frac{6}{\text{sec}}$) to calculate the angular velocity of the orientation.The Myo armband first placed on a level tabletop and then was slid along a straight line for a distance of one meter. The noise in data was filtered to adjust the initial and final velocities to zero. The velocity was obtained by integrating ac-₂₆₇ celerometer measurements once, and the position was obtained by integrating the velocity. The sensor moved a distance of one meter, the estimated distance obtained by double integration. Data from the game engine was registered based on the Unity game development platform frame-

 rates (60fps) and transferred to the computer's hardware in excel format. The average time re- $_{271}$ quired to perform each movement was found to be around 500 (milliseconds) to estimate signal

features (Lang et al. 2013).

4.2. Subjects

Subjects

Subjects

Subjects

Hifty-two purticipants were recruited for the study (forty with MS and twelve MS-free).

Hifty-two purticipants were recruited from analysis as five (from male, one fermale) could

omplete th Fifty-two participants were recruited for the study (forty with MS and twelve MS-free). Twelve MS patient's data were excluded from analysis as five (four male, one female) could not complete the therapy sessions, and seven patients decided to leave the study. Forty sub-₂₇₇ ject's data were used for decision-making and prediction; twenty-eight MS casualties and twelve healthy with age ranged [28,62]. Table 4 summarises the demographic and clinical characteris-₂₇₉ tics of the forty MS participants at the time of admission. The patients diagnosed with relapsing- remitting MS, secondary progressive MS (47%), and primary progressive MS (53%). The institu- tion (ARU) Research Ethics committee- a UK Research Integrity Office (UKRIO), approved the research. The research was performed following the Belmont principle. Informed consent was obtained from all participants or their legal guardians with a complete explanation of the research process. The explicit motor learning conditions were implemented with guidance to complete the motor tasks and learn them. It relies on maintaining attention, processing information, reflecting and maintain awareness, and working memory capacity. Participants were interviewed following each rehabilitation session to receive feedback. 83% of the MS participants had the previous experience of undertaking conventional physical and occupational therapy with discontinuation due to the lack of motivation. The eligibility criteria for MS patient's participation in the study are; slow primary and secondary progressive MS, at least 20**◦** active wrist extension, 10**◦** active ²⁹¹ finger extension, and minimal sensory or cognitive deficits (C. S. Dunham, J. R. Dunham, and Goldenson 1978).

 The healthy subjects' data also were used to train the algorithm accompanying MS casualties to measure the joint's range of motion (ROM) and hand gestures. Joint flexibility at a joint is mea- sured by the number of degrees from the starting position of a segment to its final position at the end of its full range of movement. The measurements were performed using a double-armed go- niometer and Kinect devices. In goniometer measurement, a stationary arm holding a protractor was placed parallel with a stationary body segment, and a movable arm moves along a movable body segment while the axis of the goniometer was placed over the joint (Hamilton 2011). Ta-

³⁰⁰ ble 3 displays the UE ROM collected from healthy subjects in degrees, using both goniometer 301 and Kinect devices. The control groups' data are also used to identify four-hand gestures through ³⁰² the Myo armband, where each movement was held fixed for five seconds, and the signals were ³⁰³ registered. The identified four gestures are; 1. Wave hand for reach or move to a target, 2. Fist ³⁰⁴ for grasp or hold on to an object (to carry virtual objects from point A to B), 3. Release an object ³⁰⁵ using the stop hand sign (spread fingers), and 4. Idle or relaxed hand gestures.

³⁰⁶ Motor Assessment Scale (MAS) with the scale of [0,6] was conducted pre-treatment, post-³⁰⁷ treatment, and four weeks follow up-treatment for upper arm function and hand movement. ³⁰⁸ Participants performed each task three-time; only the best performance was recorded. Two therapists supervised and obtained the participant's MAS score. If the patient could not complete 310 any part of a section, the score is zero, and if they complete it, the score is six.

³¹¹ The tasks in games are designed to incorporate the major joints of the upper body and represent 312 functional tasks that are feasible yet challenging enough to engage UE and reveal motor deficits. 313 The game difficulty, the score, and critical timing to finish each task were set initially through ³¹⁴ the game menu, which was consistent for all participants.

- Eincer devices. The control groups' data are also used to identify four-hand gestures through
fyo amband, where each movement was held fixed for five seconds, and the signals were
rered. The identified four gestures are: ³¹⁵ We performed supervised ML by supplying a known set of observations of input data (predictors) 316 and known responses by using observations to train a model that generates predicted responses 317 for new input data. As previously mentioned, the control subject's data were used to classify ³¹⁸ hand and arm patterns while players engage with the program. These data were utilized as a 319 reference to compare the patterns generated by casualties pre and post-treatment and feature ³²⁰ identification. The control subjects' within and between-session coefficients of variation were assessed $(\frac{\sigma}{\mu})$ while associated with virtual objects and completed the tasks in the games. The ³²² repeatability coefficient represents the absolute difference between two repeated test results with 323 a probability of 95%. The dispersion of the variables was ranged [42-89]%, indicating excellent ³²⁴ repeatability. Thus, control subjects' data was adequate to identify various patterns of hand and ³²⁵ arm movements and compare data. The reliability of quantitative functional tests in patients with 326 MS was considered to be varied by <20% of individual mean scores on repeated testing (Schwid 327 et al. 2002), to determine the range of measurement variability when patients are clinically stable.
	- 16

Table 3: Kinematic data of Upper Extremity average ROM collected from healthy subjects.

5. Results

5.1. Data Collection and Analysis

Drata Collection and Analysis

o develop a control strategy and prediction model of the rehabilitation, ML, was executed

that at were collected from Mya armband and Kinect devices while the players were the

ing with the To develop a control strategy and prediction model of the rehabilitation, ML was executed 331 in that data were collected from Myo armband and Kinect devices while the players were in- teracting with the game setting. The features considered for the SVM classification are as fol- lows: Shoulder joint orientation (flexion/abduction), Elbow joint orientation and position (flex- ion/extension), and Wrist joint orientation and position (flexion/extension) from Kinect (30fps), Forearm pronation/supination (50Hz), Hand gestures and data from 8-EMG Electrodes (200Hz) from Myo armband. The UE joint orientation from the control groups helped identify the upper ³³⁷ limb patterns while interacting with the game. The application of SVM classification was used to classify arm patterns pre-treatment, post-treatment, and four-month follow-up to investigate whether SVMs could be used to predict UE improvement following rehabilitation. A total of twelve parameters (six-upper arm function and six-hand movement) scores give a score between 341 [0, 72], where 72 points represent a well-aligned upper arm and hand function and full UE range 342 of motion. Three sets of data were created taken from each subject for the analysis resulting in a total data size of 105.

344 The MS people were labeled +1, and -1 demonstrated healthy people for arrangement and re- lapse ML utilizing lower arm developments and MAS scores. The region of the receiver operat- ing characteristics (ROC) curve for single highlights was utilized to get a quantitative proportion of individual element distinguishableness. ROC zones were numerically approximated utilizing the trapezoidal guideline, where more considerable qualities inferred better straight component distinguishableness. The exhibition of the ML models at recognizing healthy subjects from MS people with restricted UE was portrayed by the affectability and particularity esteems. The ROC ³⁵¹ bend territory is evaluated by utilizing the trapezoidal guideline for approximating zones under bends. The order result of the ML was additionally contrasted and linear discriminant investiga-353 tion (LDA).

 As mentioned in the previous section, data were recorded as the participants interacted with fifteen mini video games. The control subjects used their dominant arm, and the MS casual-

orcarm, band position, and the average MAS score of band and upper arm (pre/post and
contri-follownp-treatment) were used for the classifications. The set of good models was
imined by a leave one out (LOO) procedure which ³⁵⁶ ties interacted with the more impaired limb. The orientation, velocity, and muscle activity of ³⁵⁷ the forearm, hand position, and the average MAS score of hand and upper arm (pre/post and ³⁵⁸ four-month-followup-treatment) were used for the classifications. The set of good models was ³⁵⁹ determined by a leave one out (LOO) procedure which is a robust metric for determining the qual-³⁶⁰ ity of the models. All models were trained using the linear, polynomial, Radial Basis Function 361 (RBF), and Sigmoid kernels. Bayesian optimization was employed to optimize the prediction. ³⁶² Features were excluded randomly from the training data set and used as a test sample. The trial ³⁶³ sets were divided into training test sets (75:25), while the test set was not included in the training. ³⁶⁴ This was repeated until all training examples were individually tested on the models. Our test results showed that SVM and KNN are both efficient algorithms with reasonable accuracy. They could successfully classify data, identify individuals, and predict the improvement following the ³⁶⁷ rehabilitation program. The mean period of completing tasks for people with MS condition was 368 approximately twice as long as the control group (75 ± 0.6 s versus 35 ± 0.76 s).

 Scatter plots and confusion matrix, and parallel coordinates plots were combined with ob-370 serving data and assessing the accuracy of the classifier. Kernel functions with ROC areas close to unity indicated high accuracy and robustness to variations of classification inception and better overall execution.

373 The accuracy of the classifier was also identified as the rate of correct classification to all data. ³⁷⁴ It was used as the primary index to illustrate the performance of the classification. Besides, sta-³⁷⁵ tistical analyses were performed to interpret the Spatiotemporal results using a t-test in SPSS 376 Statistics 24. The test made a statistically significant difference with the mean standard error of 377 MSE=0.17.

378 Spatiotemporal data used were; movement duration, hand trajectories, velocities, the linear dis-³⁷⁹ tance between the start and endpoint, and the shoulder position during forward and upward reach-³⁸⁰ ing. Data were collected, although some participants experienced difficulties in completing some ³⁸¹ tasks, including wrist flexion, elbow extension/supination, and excessive shoulder elevation in ³⁸² forwarding/upward reaching and grasp. Participants with MS performed tasks over a higher du-383 ration than MS-free subjects 49.7% with P < 0.001. Furthermore, MS participants demonstrated

in terms of frequency of direction changes 4.2(H+) vs. 3.1(H+). The Spatiodemperal data
ning pre-and pot-treatment showed a significant improvement in real-world motor-ability
at a least four weeks post-treatment, and 1.6 reduced movement stability in terms of adjustment 5.5 vs. 1.8(mm) and a less smooth move- ment in terms of frequency of direction changes 4.2(Hz) vs. 3.1(Hz). The Spatiotemporal data comparing pre and post-treatment showed a significant improvement in real-world motor abil-³⁸⁷ ity for at least four weeks post-treatment, and 1.65s improved movement speed. The maximum speed progressed to 1.7m/s, and the timing of maximum velocity increased by 32%. Table 5 illustrates twenty-eight participant's MAS average upper arm function and its standard deviation (SD). <20% change on the test result is considered the threshold that reliably indicates an ac-³⁹¹ tual change in function for an individual with MS. An unbiased training set was derived using a leave-one-out cross-validation method in that; an arbitrary example was excluded from the train- ing data and used as a test sample. This was repeated until all training samples were individually tested on the SVM. The average accuracy is the leave-one-out accuracy and is a robust metric for determining the quality of the SVM model.

 The features included in the classification were; Wrist, Elbow, Forearm's joint/segment Orien- tation, Velocity, Position, EMGs, MAS Scores, and Time. MS patients were identified by -1 and healthy subjects as +1. The five-fold cross-validation and holdout validation classifier were used to overcome data overfitting. Five-fold cross-validation partitions the data into five disjoint sets and trains the model using the out-of-fold observations. It assesses the model's performance using in-fold data by studying the average test error overall folds. Table 6 lists the standardized kernel functions and RMSEs in that the overall error rate was used as the guiding criterion for performance. The Kernels scale mode was set to both automatic and manual. The automatic scale uses a heuristic procedure to select the initial kernel parameters, which the initial values were specified in manual one.

 The automatic Gaussian kernel function has the best fit with RMSE=0.056. The manual Gaus- sian kernel function with the box constraint=1.483, epsilon=0.148 and scale=21 resulted in the RMSE=0.079. The features were reduced to Elbow, Forearm's joint/segment Orientation, Veloc- ity, EMGs, MAS Scores, and Time in that the classification was improved. The residuals plot was investigated to check model performance considering the difference between the predicted and ⁴¹¹ true responses. The kernel functions determine the correlation in the response as a function of

ions, and Bayesian Optimization to build SVM and compared the outcomes. The Bayesian izzation was employed to optimize the SVM classifier fit.

fold cross-validation and 25% holdout validation were utilized to partition th ⁴¹² the distance between the predictor values. We adopted linear, polynomial, RBF, Sigmoid kernel 413 functions, and Bayesian Optimization to build SVM and compared the outcomes. The Bayesian ⁴¹⁴ Optimization was employed to optimize the SVM classifier fit. ⁴¹⁵ Five-fold cross-validation and 25% holdout validation were utilized to partition the data set into ⁴¹⁶ folds, estimate the training data accuracy, and avoid over-fitting. The results indicated that the ⁴¹⁷ SVM could successfully identify individuals with MS from the healthy using the data collected ⁴¹⁸ from sensors and predict rehabilitation outcomes. The out-of-sample misclassification rate was ⁴¹⁹ 5% which indicated a reasonable classification. 25% holdout validation classifier was used in ⁴²⁰ that medium Gaussian radial basis function had the highest accuracy, 83.8% as displayed in 421 Table 7. Figure 3 shows the 3D model of the objective functions that are used to perform op-⁴²² timization. The optimal values of the decision variables result in the best possible value of the ⁴²³ objective function to find an input that results in the minimum/maximum cost of a given objec-⁴²⁴ tive function. Bayesian optimization offers a principled technique based on the Bayes Theorem ⁴²⁵ to supervise a search of a global optimization problem that is efficient and effective. It starts 426 by building a probabilistic model of the objective function, known as the surrogate function, ⁴²⁷ then searched efficiently with an acquisition function before applicant samples are collected to 428 evaluate the actual objective function. As illustrated in figure thirty, Objective Evaluations and 429 Sigmoid kernel function were adopted to train the classifier. The results showed a good fit and 430 low cross-validation loss of the pairs using Bayesian Optimization. The characterization takes a ⁴³¹ shot at areas of focuses from a Gaussian blend model with base focuses produced arbitrarily and ⁴³² freely.

433 **6. Discussion**

434 Multiple Sclerosis (MS) is a degenerative neurological issue influencing casualties freedom 435 of everyday life; hence, causing dependency on others (Taylor and Griffin 2015). Studies have ⁴³⁶ shown the advantage of continuous rehabilitation and its role in a patient's brain plasticity and ⁴³⁷ recuperation at various phases of the disease (S. Thomas et al. 2017). Plasticity is the capac-⁴³⁸ ity of the sensory system to adjust to the ever-changing states of the environment, experienced

Table 6: SVM Regression Learner models using Cross-validation and Holdout-Validation.

Table 7: Binary Classification using SVM and KNN algorithms.

Area Under the curve (AUC);

Receiver Operating Characteristic (ROC).

Figure 3: Optimized SVM classifier fit using Bayesian Optimization. The output and input pairs' of objective function model.

Superior and the start of t during improvement and learning (Prosperini et al. 2015). This study integrated the Human Ma- chine Interface platform to facilitate exercise and rehabilitation programs for MS casualties to reduce the limitations resulting from the neurological deficit. ML was used for data exploration ⁴⁴² and to gain the experience required to develop predictive models, classify patterns, and reduce 443 dimensionality in human UE movement analysis. The application of ML methods to human posture and movement classification was studied to assess the movement quality in response 445 to the intervention. The study demonstrated the application of KNN and SVM to predict the outcome of the home-based neuro-rehabilitation video game and its effectiveness in supporting 447 clinical decision-making. KNN and SVM have the advantage of high accuracy, good theoretical guarantees regarding overfitting. SVM performs well where base feature and kinematic data are interconnected and are not linearly separable. KNN is a high variance classifier and demonstrates robustness towards noisy training data.

⁴⁵¹ NN and SVM both showed having the ability to recognize complex and noisy patterns available ⁴⁵² in UE (upper extremity) data. SVM and KNN are both efficient while using cross-validation and

 holdout validation. The algorithm could identify individuals with a motor deficit from healthy participants, recognize the patterns created by joint kinematics (both healthy and MS casualties) and create a regression model that predicts the rehabilitation advantages post-treatment. The prediction accuracy was in the range of [61% to 91.7%], where Gaussian KNN exceeded

other kernel functions. Bayesian optimization resulted in an excellent classifier fit.

eignats, recognize the patterns created by joint kinematics (both healthy and MS casualtics)
reative a regression model that predicts the rehabilitation advantages post-treatment.
rerediction accuracy was in the range of [Spatiotemporal control and motor assessment scale used to assess the effectiveness of the rehabilitation program. Our findings suggest that ML has a great potential to be used for decision- making in neuro-rehabilitation programs. It is suggested that more accurate targeting of rehabili-⁴⁶¹ tation programs could result in the more efficient and effective use of health care resources, which results in the reduction of health care costs. The outcomes also illustrate that repetitive move- ment of the affected UE enables neuro-plasticity in MS casualties. Real-time graphical feedback provided in the invention offered a suitable platform for self-assessment, suggested prescriptive input about what to do following a mistake/dysfunction, thus, reducing the cognitive load. The participants agreed that the video game platform increased their motivation and retention com-467 pared to their experiences with conventional therapies. The theory and mechanism behind the improvements in outcomes following the interaction with the rehabilitation program indicated the adequacy of computer games and visual-input practices in improving parity in MS patients.

7. Conclusion

⁴⁷¹ With the expanded use of standardized information systems in many parts of the health sys- tem, there is great potential for enhanced use of sophisticated computer modeling and statistical 473 analysis techniques to inform clinical decision-making and health system planning (M. Zhu et al. 474 2007). Data selection to train SVM and KNN models requires randomly selected large samples to achieve satisfactory models. However, many training data will raise the computation cost for model training and data testing. Thus, it is crucial to reduce the training dataset without degrad-⁴⁷⁷ ing the final classification result. The results showed that to train a multiclass support vector machine (SVM), a small training set is sufficient since an initial model is improved iteratively 479 and incrementally, which complies with the study of (Zhang et al. 2011). It is important to

three sample sizes. Our models were evaluated and validated against the elassification ap-
thin literature; (Fleury, Vacher, and Noury 2010; Keethi and Lin 2003; Lau, Tong, and
hin literature; (Fleury, Vacher, and Noury 2 choose the right parameters to achieve testing accuracy rather than using conventional models 481 with large sample sizes. Our models were evaluated and validated against the classification ap- proach in literature; (Fleury, Vacher, and Noury 2010; Keerthi and Lin 2003; Lau, Tong, and H. Zhu 2009). The classification derived from the small training set improved computational efficiency by 28%.

 The use of the proposed rehabilitation study could be considered experimental in MS people. Thus further research with rigorous methodology and reliable outcome measures could be bene- ficial to provide higher-level support of the effectiveness of the intervention for MS people. The future work will focus on using the classification data to achieve a dynamical coupling between a player and a prosthetic armor artificial limb. This could lead to an expansion of ML functionality

490 in developing advanced online training schemes to support the long-term operation.

Funding

This research received no external funding.

Conflicts of Interest

The authors declare no conflict of interest.

References

Begg, Rezaul K, Marimuthu Palaniswami, and Brendan Owen (2005). "Support vector machines for automated gait

classification". In: *IEEE transactions on Biomedical Engineering* 52.5, pp. 828–838.

- Benedict, Ralph HB et al. (2005). "Predicting quality of life in multiple sclerosis: accounting for physical disability,
- fatigue, cognition, mood disorder, personality, and behavior change". In: *Journal of the neurological sciences* 231.1- 2, pp. 29–34.
- Bettger, Janet A Prvu and Margaret G Stineman (2007). "Effectiveness of multidisciplinary rehabilitation services in postacute care: state-of-the-science. A review". In: *Archives of physical medicine and rehabilitation* 88.11, pp. 1526– 1534.
- Boser, Bernhard E, Isabelle M Guyon, and Vladimir N Vapnik (1992). "A training algorithm for optimal margin classi-
- fiers". In: *Proceedings of the fifth annual workshop on Computational learning theory*. ACM, pp. 144–152.

- Butler, Erin E et al. (2010). "Temporal–spatial parameters of the upper limb during a reach & grasp cycle for children".
- In: *Gait* & *posture* 32.3, pp. 301–306.
- Corbetta, Davide et al. (2015). "Constraint-induced movement therapy for upper extremities in people with stroke". In: *The Cochrane Library*.
- Cortes, Corinna and Vladimir Vapnik (1995). "Support-vector networks". In: *Machine learning* 20.3, pp. 273–297.
- Côté-Allard, Ulysse et al. (2018). "Deep Learning for Electromyographic Hand Gesture Signal Classification by Lever-

aging Transfer Learning". In: *arXiv preprint arXiv:1801.07756*.

- Cowan, J, IE Ormerod, and P Rudge (1990). "Hemiparetic multiple sclerosis." In: *Journal of Neurology, Neurosurgery* & *Psychiatry* 53.8, pp. 675–680.
- Derie, Rud et al. (2020). "Tibial acceleration-based prediction of maximal vertical loading rate during overground run-ning: a machine learning approach". In: *Frontiers in bioengineering and biotechnology* 8, p. 33.
- E Gati & parame 32.3, pp. 301–306.
I.L. Dowsise et . (2013). "Constraint-induced movement themps for upper extensities in people with shock". In:

Jour Dowsis et . (2013). "Constraint-induced movement themps for upper exte Dunham, Charlis S, Jerome R Dunham, and Robert M Goldenson (1978). *Disability and rehabilitation handbook*. McGraw-Hill.
- Esfahlani, Shabnam Sadeghi, Javaid Butt, and Hassan Shirvani (2019). "Fusion of Artificial Intelligence in Neuro-Rehabilitation Video Games". In: *IEEE Access* 7, pp. 102617–102627.
- Esfahlani, Shabnam Sadeghi, Bogdan Muresan, et al. (2018). "Validity of the Kinect and Myo armband in a serious game
- for assessing upper limb movement". In: *Entertainment Computing*.
- Esfahlani, Shabnam Sadeghi, Tommy Thompson, et al. (2018). "ReHabgame: A non-immersive virtual reality rehabili-tation system with applications in neuroscience". In: *Heliyon* 27, e00526.
- Fleury, Anthony, Michel Vacher, and Norbert Noury (2010). "SVM-based multimodal classification of activities of daily
- living in health smart homes: sensors, algorithms, and first experimental results". In: *IEEE transactions on informa-tion technology in biomedicine* 14.2, pp. 274–283.
- Giusti, Andrea et al. (2006). "An analysis of the feasibility of home rehabilitation among elderly people with proximal
- femoral fractures". In: *Archives of physical medicine and rehabilitation* 87.6, pp. 826–831.
- Hamilton, Nancy P (2011). *Kinesiology: Scientific basis of human motion*. Brown & Benchmark.
- Hastie, Trevor, Robert Tibshirani, and Jerome Friedman (2009). *The elements of statistical learning: data mining, infer-ence, and prediction*. Springer Science & Business Media.
- Heisele, Bernd, Purdy Ho, and Tomaso Poggio (2001). "Face recognition with support vector machines: Global ver-
- sus component-based approach". In: *Computer Vision, 2001. ICCV 2001. Proceedings. Eighth IEEE International Conference on*. Vol. 2. IEEE, pp. 688–694.
- Hosomi, Masashi et al. (2012). "A modified method for constraint-induced movement therapy: a supervised self-training
- protocol". In: *Journal of Stroke and Cerebrovascular Diseases* 21.8, pp. 767–775.
- Joachims, Thorsten (1998). "Text categorization with support vector machines: Learning with many relevant features".
- In: *European conference on machine learning*. Springer, pp. 137–142.
	-

- Jonsdottir, Johanna et al. (2018). "Serious games for arm rehabilitation of persons with multiple sclerosis. A randomized controlled pilot study". In: *Multiple sclerosis and related disorders* 19, pp. 25–29.
- Keerthi, S Sathiya and Chih-Jen Lin (2003). "Asymptotic behaviors of support vector machines with Gaussian kernel". In: *Neural computation* 15.7, pp. 1667–1689.
- Kwakkel, G et al. (2019). "Standardized measurement of quality of upper limb movement after stroke: consensus-based
- core recommendations from the Second Stroke Recovery and Rehabilitation Roundtable". In: *Neurorehabilitation and neural repair* 33.11, pp. 951–958.
- Kwakkel, Gert et al. (2015). "Constraint-induced movement therapy after stroke". In: *The Lancet Neurology* 14.2, pp. 224–234.
- Lang, Catherine E et al. (2013). "Assessment of upper extremity impairment, function, and activity after stroke: founda-tions for clinical decision making". In: *Journal of Hand Therapy* 26.2, pp. 104–115.
- Lau, Hong-yin, Kai-yu Tong, and Hailong Zhu (2009). "Support vector machine for classification of walking conditions of persons after stroke with dropped foot". In: *Human movement science* 28.4, pp. 504–514.
- Levinger, Pazit et al. (2007). "Support Vector Machines for detecting recovery from knee replacement surgery using
- quantitative gait measures". In: *Engineering in Medicine and Biology Society, 2007. EMBS 2007. 29th Annual International Conference of the IEEE*. IEEE, pp. 4875–4878.
- Long, Yi et al. (2016). "PSO-SVM-based online locomotion mode identification for rehabilitation robotic exoskeletons". In: *Sensors* 16.9, p. 1408.
- Mark, Victor W et al. (2008). "Constraint-Induced Movement therapy can improve hemiparetic progressive multiple sclerosis. Preliminary findings". In: *Multiple Sclerosis Journal* 14.7, pp. 992–994.
- mention plus usedy". In: *Multiple schemati and related disorders* 19, pp. 25–29.

J. S Solony and Childrels Lain (2003) "Asyngotic behavior of support vector mechanics with Giovsina locoed".

J. Sochlays and Childrels Nazmi, Nurhazimah et al. (2016). "A review of classification techniques of EMG signals during isotonic and isometric contractions". In: *Sensors* 16.8, p. 1304.
- Oskoei, Mohammadreza Asghari, Huosheng Hu, et al. (2008). "Support vector machine-based classification scheme for myoelectric control applied to upper limb." In: *IEEE Trans. Biomed. Engineering* 55.8, pp. 1956–1965.
- Page, Stephen J et al. (2008). "Modified constraint-induced therapy in chronic stroke: results of a single-blinded random-ized controlled trial". In: *Physical therapy* 88.3, pp. 333–340.
- Phinyomark, Angkoon, Rami N Khushaba, and Erik Scheme (2018). "Feature Extraction and Selection for Myoelectric Control Based on Wearable EMG Sensors". In: *Sensors* 18.5, p. 1615.
- Prosperini, Luca et al. (2015). "Functional and structural brain plasticity enhanced by motor and cognitive rehabilitation in multiple sclerosis". In: *Neural plasticity* 2015.
- Schwid, SR et al. (2002). "Quantitative functional measures in MS: what is a reliable change?" In: *Neurology* 58.8, pp. 1294–1296.
- Steinman, Lawrence (1996). "Multiple sclerosis: a coordinated immunological attack against myelin in the central ner-
- vous system". In: *Cell* 85.3, pp. 299–302.
-

- Stetter, Bernd J et al. (2020). "A machine learning and wearable sensor based approach to estimate external knee flexion
- and adduction moments during various locomotion tasks". In: *Frontiers in bioengineering and biotechnology* 8.
- Tannous, Halim Elie (2018). "Interactive and connected rehabilitation systems for e-health". PhD thesis. Université de Technologie de Compiègne.
- at adduction mements during various locemesion task.". In: *Fremciers* in *biomegineering and hidechomology* 8.

In, Halm Eliz (2018). "Interactive and connected relativitionion systems for e-health." Full these Universit Taub, Edward, Jean E Crago, and Gitendra Uswatte (1998). "Constraint-induced movement therapy: A new approach to treatment in physical rehabilitation." In: *Rehabilitation Psychology* 43.2, p. 152.
- Taub, Edward, Gitendra Uswatte, et al. (2006). "A placebo-controlled trial of constraint-induced movement therapy for upper extremity after stroke". In: *Stroke* 37.4, pp. 1045–1049.
- Taylor, Matthew JD and Murray Griffin (2015). "The use of gaming technology for rehabilitation in people with multiple sclerosis". In: *Multiple Sclerosis Journal* 21.4, pp. 355–371.
- Thomas, Sarah et al. (2017). "Mii-vitaliSe: a pilot randomised controlled trial of a home gaming system (Nintendo Wii) to increase activity levels, vitality and well-being in people with multiple sclerosis". In: *BMJ open* 7.9, e016966.
- Wolf, Steven L et al. (2006). "Effect of constraint-induced movement therapy on upper extremity function 3 to 9 months after stroke: the EXCITE randomized clinical trial". In: *Jama* 296.17, pp. 2095–2104.
- Wu, Ge et al. (2005). "ISB recommendation on definitions of joint coordinate systems of various joints for the reporting
- of human joint motion—Part II: shoulder, elbow, wrist and hand". In: *Journal of biomechanics* 38.5, pp. 981–992.
- Yen, Jyh-Geng et al. (2005). "Effectiveness of modified constraint-induced movement therapy on upper limb function in stroke subjects". In: *Acta Neurol Taiwan* 14.1, pp. 16–20.
- Zhang, Shumei et al. (2011). "Optimal model selection for posture recognition in home-based healthcare". In: *Interna-*
- *tional Journal of Machine Learning and Cybernetics* 2.1, pp. 1–14.
- Zhu, Mu et al. (2007). "The K-nearest neighbor algorithm predicted rehabilitation potential better than current Clinical
- Assessment Protocol". In: *Journal of clinical epidemiology* 60.10, pp. 1015–1021.

⁵⁹⁶ Abbreviations

⁵⁹⁹ Data Availability

600 Samples of data are available at; Supplementary Material.

29

You are currently signed in as S. S. Esfahlani **https://orcid.org/0000-0002-1443-0330**

Journal Partners

Highlights: Machine Learning in Neuro-Rehabilitation Video Game

- 1. Using Supervised learning algorithms; Support Vector Machines (SVM) and K-Nearest Neighbours (KNN), for motor function classification and regression analysis
- 2. To achieve clinical decision-making and prediction of motor function of participant's with Multiple sclerosis
- 3. Mini video games was developed for upper limb rehabilitation

1. Using Supervised learning algorithms; Support Vector Machines (SVM)
and K-Nourse Kraghbours (KNN), for motor function dussification and
regression analysis
2. To achieve educated decision-making and prediction of motor 1

Conceptualization, K.S.E;

Methodology, S.S.E.;

Software, S.S.E.;

Validation, S.S.E., and H.S.;

Formal analysis, S.S.E., J.B.;

Investigation, S.S.E., J.B. and H.S.;

Resources, I.M.;

Data curating, S.S.E.;

Writing original draft preparation, S.S.E.;

Writing review and editing, S.S.E. and H.S.;

Visualization, S.S.E. and H.S.;

Supervision, S.S.E.;

Journal Pre-Journal Pre-proper Project administration, S.S.E., J.B., K.S.E.

Declaration of interests

 \boxtimes The authors declare that they have no known competing financial interests or personal relationships that could have appeared to influence the work reported in this paper.

and a character that they have no known competing financial interests or personal relationships
are appeared to influence the work reported in this paper.
Its declare the following financial interests/personal relationship ☐The authors declare the following financial interests/personal relationships which may be considered as potential competing interests: